


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR 17-1-PS PN

S.N.: 8101

Date: JUNE 23 1997

- a) Output signal amplitude: 0 TO ± 1000 VOLTS
- b) Pulse width: ($R_L \geq 200 \Omega$)
0.2 TO 5.0 μ S
- c) Rise time: ≤ 50 NS
- d) Fall time: ≤ 50 NS
- e) PRF: 0 TO 1 KHz.
- f) Jitter, stability: OK
- g) Prime power: 100 $\pm 10\%$
50-60 Hz.



$R_L = 10K$
500 V/DIV
500 NS/DIV
PRF = 1 KHz

