


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P-EW*

S.N.: *B124*

Date: *JUN 22 1997*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 VDC
600 mA (max)

40db ATTN

10 VOLTS/DIV

P.R.F = 10 MHz