


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2P-EW*

S.N.: *B125*

Date: *JUN 22 1997*



- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*0.2 TO 2.0 NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*724 VDC*

*40 dB ATTN*

*-10 VOLTS/PIV*

*PRF = 10 MHz*

*600 mA (MAX)*

*[Signature]*