


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ADRF-4-C-P*

S.N.: *3149*

Date: *Aug 20 1997*



- a) Output signal amplitude: *0 TO +400 V*
- b) Pulse width: *CTD $R_c = 50 \Omega$
0.1 TO 10 μ S*
- c) Rise time: *≤ 5 NS*
- d) Fall time: *≤ 5 NS*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 V
50-60 Hz.*

60 dB ATTN
100 VOLTS/DIV
HORIZ: 20 NS/DIV
PRF = 1 KHz