


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-E3 P*

S.N.: *8151*

Date: *Aug 27 1997*



- a) Output signal amplitude:  
*0 TO +100 V (+0.50V)*
- b) Pulse width:  
*20 TO 200 NS*
- c) Rise time:  
*≤ 0.5 NS.*
- d) Fall time:  
*≤ 1.0 NS*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 VDC*  
*400 mA (max)*

*60 db ATTEN*

*100 VOLTS/DIV*

*TOP* *50 NS/DIV*

*MID* *5 NS/DIV*

*BOI* *0.5 NS/DIV (RESETTING)*

*[Signature]*