


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-9B-C-P-0P1.*

S.N.: *8203*

Date: *NOV 4 1997*



- a) Output signal amplitude: *0 TO +1.0 VPP*
- b) Pulse width: *20 TO 200NS*
- c) Rise time: *≤ 0.5NS*
- d) Fall time: *≤ 0.5NS*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240V*
50-60 Hz,

TOP MANIPULATE OUT TO
50 db, 10 NS/DIV

MID AS TOP BUT *1 NS/DIV*
(RISE TIME)

BOI *OUT TO 30 db,*
(IN 5819 DIODE LOAD)..
PRF ≈ 50 KHz,
10 NS/DIV.