


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-HV3-C-P*
 S.N.: *8218*
 Date: *NOV 6 1997*



- a) Output signal amplitude:
0 TO + 100 VOLTS
- b) Pulse width:
4 - 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
O/K
- g) Prime power:
120/240V
50-60Hz

60 dB ATTN.: 100 VOLTS/DIV

TOP 10 NS/DIV

*BOT 200 PS/DIV
(RISE TIME)*

PRF = 25 KHz

[Signature]