


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVO-6D-C-N-1M-0P1*

S.N.: *8232*

Date: *OCT 15 1997*



*AVO-6D-T ME out to 50Ω  
50V/DIV, 200 NS/DIV*



*AVO-6D-T ME out to 50Ω  
2V/DIV - 2 AMP/DIV  
200 NS/DIV  
1N5819 TEST DIODE*

- a) Output signal amplitude:  
*0 TO - 10 AMPS*
- b) Pulse width:  
*50NS TO 5.00μS  
(1% MAX DUTY CYCLE)*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF: *0 TO 10 KHz  
(1% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*0.1%*
- g) Prime power: *120/240V  
50-60 Hz*