


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV14-5-1-C-P

S.N.: 8243

Date: SEPT 22 1997



- a) Output signal amplitude: 0 TO +10V0V3
- b) Pulse width: ≤ 130 PS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: O.K.
- g) Prime power: 120/240V
50-60 Hz

[Signature]

40 db ATTEN
 10 V0V3/DIV
 TR 100 ps/DIV
 BOT 1.0 ns/DIV
 PRF = 200 kHz