


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVP-AV-1C-P-05*

S.N.: *8244*

Date: *SEPT 22 1997*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz

[Signature]

TDR 1 NS/DIV

BOI 100 PS/DIV (RISE TIME)

40 dB ATTN

10 VOLTS/DIV

PRF = 100 MHz