


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-6D-C-PN-07-05-0P1


S.N.: 8298

Date: JAN 27 1998



- a) Output signal amplitude:
0 TO ±10 AMP
- b) Pulse width:
50 NS TO 5.0 μS
(1% MAX DUTY CYCLE)
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF: 0 TO 10 KHz
(1% MAX DUTY CYCLE)
- f) Jitter, stability:
OK

⊕ MAINFRAME OUT TO 50 Ω (pos)



- g) Prime power: 120/240 V
50-60 Hz

⊕ AVO-6D-T-05 OUT TO
1N5819 IN SERIES WITH
1 Ω.

