


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMH-1-P*

S.N.: *8305*

Date: *DEC 1 1997*



- a) Output signal amplitude: *0 TO +10V*
- b) Pulse width: *≤ 200 ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24 VDC*  
*400 mA DC*

*40 dB ATTEN*  
*10 V/VTB 1DIV*  
*PRF = 5 MHz*

*WJ*