


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-2-C-N-PN-M-CA-EW-EO*

S.N.: *8316*

Date: *MAR 16 1998*



- a) Output signal amplitude:
0 TO 120 VOLTS
- b) Pulse width: *($V_L = 50V$)*
A) 0.4 TO 8.0NS
B) 2.0 TO 100NS
- c) Rise time:
≤ 200NS
- d) Fall time:
≤ 300PS
- e) PRF:
0 TO 100KHz
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz.

40 db ATTEN
- 10 VOLTS/DIV

TOP *1 NS/DIV*
(5 NS RANGE)

BOT *10 NS/DIV*
100 NS/RANGE