


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-4B-N-PS-US2

S.N.: 8350

Date: APRIL 7 1998



- a) Output signal amplitude: ≈ -1000 VOLTS
- b) Pulse width: 50-130 pps
3 ns (AT 900 V)
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 10 KHZ
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 HZ

$R_L = 50 \Omega$ (C=optd)

70 dB ATTEN

- - 320 VOLTS 10 V

PRF = 50 KHZ