


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-4B-N-PS-VS2

S.N.: 8351

Date: APRIL 15 1998



- a) Output signal amplitude: ≥ -1000 VOLTS (TO 50Ω / 130pF)
 - b) Pulse width: 3 NS (AT 900V)
 - c) Rise time: NA
 - d) Fall time: NA
 - e) PRF: 0 TO 10 KHz
 - f) Jitter, stability: OK
 - g) Prime power: 120/240 V
50-60 Hz
- [Signature]*

$R_L = 50\Omega$ ($C_L = 0\text{pF}$)

70 dB ATTEN

$\therefore 320$ VOLTS/DIV

PRF = 5 KHz