


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-PS-P-CBNL1*

S.N.: *8359*

Date: *FEB 13 1998*



- a) Output signal amplitude:
0 TO 10 V RMS
- b) Pulse width:
1 TO 5 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 350 PS
- e) PRF: *0 TO 2.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50/60 Hz

TR 1 NS/DIV
BOI 200 PS/DIV
PRF = 1.6 MHz
40 dB ATTN
10 V RMS/DIV