


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM-3-N*

S.N.: *8361*

Date: *APRIL 6 1998*



- a) Output signal amplitude: *0 TO -15 V.4*
- b) Pulse width: *2 TO 15 NS*
- c) Rise time: *≤ 150 PS*
- d) Fall time: *≤ 600 PS*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+28 VDC*  
*500 mA*  
*MAX.*

*40 db ATTEN*  
*10 VOLTS/DIV*  
*PRF = 10 MHz*