


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANRF-4-CP*

S.N.: *8366*

Date: *JAN 12 1998*



- a) Output signal amplitude:  
*0 TO +400 V*
- b) Pulse width: ( $R_L = 50 \Omega$ )  
*0.1 TO 1.0  $\mu$ S*
- c) Rise time:  
 *$\leq 5$  ns*
- d) Fall time:  
 *$\leq 5$  ns*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *120/240 V*  
*50 - 60 Hz.*

*60 dB ATTEN*  
*100 V/DIV*  
*20 NS/DIV*  
*PRF = 1 KHz.*

*[Signature]*