


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MH-141-C-P-T1*

S.N.: *8396*

Date: *MAR 27 98*



- a) Output signal amplitude: *0 TO +100 VOLTS*
- b) Pulse width: *600 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60 Hz*

*60 db ATTN*

*100 VOLTS/DIV*

*PRF = 100 KHz*

A large, stylized handwritten signature or scribble, possibly reading 'OK' or similar, located at the bottom right of the page.