


PULSE GENERATOR
PERFORMANCE CHECK

Model: *M-107C-PS-P-GA7LI*

S.N.: *3406*

Date: *MAY 6 1988*



- a) Output signal amplitude: *0 TO +20 AMPS*
(TO 0 TO +100 VOLTS)
- b) Pulse width: *0.5 TO 20 μ S.*
- c) Rise time: *≤ 100 NS*
- d) Fall time: *≤ 100 NS*
- e) PRF: *≤ 1 HZ*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz.

TOP *20V/DIV, 1 μ S/DIV*
R_c = 5 Ω

BOT *50V/DIV, 20 μ S/DIV*
R_c = 5 Ω , DOUBLE PULSE