


PULSE GENERATOR
PERFORMANCE CHECK

Model: AURL-5-P5

S.N.: B411

Date: JUNE 14 1998



- a) Output signal amplitude: -200 VOLTS (RL \geq 10 Ω)
- b) Pulse width: 20 NS TO 1.0 MS
- c) Rise time: \leq 10 NS
- d) Fall time: \leq 10 NS
- e) PRF: 0 TO 1 KHz.
- f) Jitter, stability: OK
- g) Prime power: 140/200V
50-60 Hz.

TOP 50 NS / DIV
(RANGE A)

BOI 50 US / DIV
(RANGE B)