


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVRH-3-PS-PN*

S.N.: *8425*

Date: *JUNE 10 1998*



- a) Output signal amplitude:  
*0 TO ±3000 V (R<sub>L</sub> ≥ 10K)*
- b) Pulse width:  
*200 NS TO 2.5 μS*
- c) Rise time:  
*≤ 100 NS*
- d) Fall time:  
*≤ 100 NS*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *120/240V*  
*50-60Hz*

