


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-6A-C-0P1B-7-M*

S.N.: *8428*


Date: *JUNE 11 1998*




- a) Output signal amplitude:  
*0 TO +10 AMP*
- b) Pulse width:  
*50 NS TO 5.0 μS*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF:  
*0 TO 20 KHz*

Ⓐ *PLW MIN CURRENT TRANSFORMER OUT, 1 AMP/DIV 50 NS/DIV*

- f) Jitter, stability:  
*OK*
- g) Prime power: *120/240V 50-60Hz*



Ⓑ *PLW MAX TOP: CURRENT TRANSFORMER OUT  
PUL: 1 μS RESISTOR VOLTAGE*



*[Signature]*