


PULSE GENERATOR  
PERFORMANCE CHECK

Model: ANI-U-1403-C →

S.N.: 8430

Date: June 18 1992



- a) Output signal amplitude:  
0 TO +100V (50.Ω)
- b) Pulse width:  
4-100 NS
- c) Rise time:  
≤ 200PS
- d) Fall time:  
≤ 4 NS
- e) PRF:  
0 TO 5 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V

60 db ATTEN

100 VOLTS/DIV

TOP 5 NS/DIV

BOT 500 ps/DIV  
(RISE TIME)

PRF = 5 KHZ.

50-60 KHZ.