


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVG-4A-7

S.N.: 8436

Date: JUNE 24 1998



- a) Output signal amplitude:
0 TO +650 V
- b) Pulse width: (50%)
≤ 5 NS (20% point)
- c) Rise time:
≤ 1.5 NS
- d) Fall time:
≤ 3 NS
- e) PRF: 0 TO 10 KHz
- f) Jitter, stability: OK
- g) Prime power: +15 VDC
450 mA MAX

60 dB ATTEN

100 V/DIV

2 NS/DIV

PRF = 10 KHz

[Signature]