


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-7A-C-P-081*

S.N.: *8506*

Date: *APRIL 29 1998*



- a) Output signal amplitude: *0 TO +700 VOLTS*
- b) Pulse width: *($R_L \geq 50 \Omega$)
50NS TO 5.0US*
- c) Rise time: *(0.5% MAX DUTY CYC)
 $\leq 50 \text{ NS}$*
- d) Fall time: *$\leq 50 \text{ NS}$*
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability: *(0.5% MAX DUTY CYCLES)
OK*
- g) Prime power: *120/240V
50-160Hz.*

$R_L = 50 \Omega$

VERT: 150 VOLTS/DIV

TOP 500 NS/DIV

BOT 100 NS/DIV.

[Signature]