


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-3-PS-PN-LUC2*

S.N.: *8510*

Date: *AUG 13 1998*



a) Output signal amplitude:
0 to ±250 V, $R_L \geq 100 \Omega$

b) Pulse width:
100 ns to 10 μ s

c) Rise time: *$\leq 50 ns$*


d) Fall time:
 $< 50 ns$

e) PRF: *0 to 25 kHz*

f) Jitter, stability: *OK*

g) Prime power: *120/240 V
50-60 Hz*

P out to $R_L = 100 \Omega$ AM.



*N_{out} ENABLE $\approx 0.5 ms$
TR₆ $\approx 1 \mu s$
(NOT CALIBRATED)*