


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 8522

Date: MAY 20 1998



- a) Output signal amplitude:
 $\geq +40$ VOLTS ✓
- b) Pulse width:
 1 TO 100 NS ✓
- c) Rise time:
 ≤ 350 PS ✓
- d) Fall time:
 ≤ 2 NS ✓
- e) PRF:
 MIN PW: 0 TO 1 MHz ✓
 MAX PW: 0 TO 20 kHz ✓
- f) Jitter, stability:
 OK ✓
- g) Prime power:
 +15 VDC ✓
 100 mA (MAX) ✓

50 dB ATTEN
∴ 32 VOLTS/DIV
PRF = 30 kHz