


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 8526

Date: MAY 20 1998



- a) Output signal amplitude:  
 $\geq +40$  VOLTS ✓
- b) Pulse width:  
1 TO 100 NS ✓
- c) Rise time:  
 $\leq 350$  PS ✓
- d) Fall time:  
 $\leq 2$  NS ✓
- e) PRF:  
MIN PW: 0 TO 1 MHz ✓  
MAX PW: 0 TO 20 kHz ✓
- f) Jitter, stability:  
OK ✓
- g) Prime power: ✓  
+15 VDC  
100 mA (MAX)

50 dB ATTEN  
 $\therefore 32$  VOLTS/DIV  
PRF = 30 kHz