


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-C-N-0P1B*

S.N.: *8537*

Date: *AUG 26 1998*



- a) Output signal amplitude:  
*0 TO -350 V TO  $R_L = 50 \Omega$*
- b) Pulse width:  
*5 TO 100 ns*
- c) Rise time:  
 *$\leq 2$  ns*
- d) Fall time:  
 *$\leq 4$  ns*
- e) PRF: *0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *120/240 V*  
*50-60 Hz*

*60 db ATTEN*  
*∴ 100 VOLTS/DIV*  
*TOP 10 ns/DIV*  
*BOI 2 ns/DIV*  
*(RISE TIME)*  
*PRF = 2 kHz*