


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-17MP-D-400A*

S.N.: *8547*

Date: *MAY 25 1998*



- a) Output signal amplitude:
 $\geq +40V$
- b) Pulse width:
1 to 100 ns
- c) Rise time:
 $\leq 350 ps$
- d) Fall time:
 $\leq 2 ns$
- e) PRF: *MAIN PW: 0 TO 1 MHz*
MIN PW: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

50 db ATTEN

32 VOLTS/DIV

PRF = 20 KHz

+15 VDC
100 mA (MAX)