


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVNC-2-C-RAY1

S.N.: 8552

Date: AUG 6 1948



- a) Output signal amplitude: ECL LEVELS
-0.9 & -1.8V (TO 50mV TO -2V)
- b) Pulse width: 50% DUTY CYCLE
- c) Rise time: ≤ 1 NS
- d) Fall time: ≤ 1 NS
- e) PRF: 10 KHZ TO 250 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60Hz

TOP OUT FOR RANGE A
 ≈ 50 MHz

POT OUT FOR TOP
1 VOLT/DIV
5 NS/DIV

