


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-U-C-PN-M-OS-INDA-OPIB

S.N.: 8564

Date: OCT 27 1998



- a) Output signal amplitude:
0 TO $\pm 50V$ (TO $50mV$)
- b) Pulse width:
2 NS TO 100 NS
- c) Rise time:
 $\leq 250 ps$
- d) Fall time:
 $\leq 2 ns$
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

50 db ATTN
2.32 V/DIV
TOP 10 NS/DIV (PONT)
MID 200 ps/DIV (PANT)
BOT 5 NS/DIV (NONT)