


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMM-2-PS-TR-OT-P*

S.N.: *8567*

Date: *NOV 26 1998*



a) Output signal amplitude:  
*0 TO +5V (50Ω)*

b) Pulse width:  
*1 TO 10 NS*

*0.3 NS*) Rise time: *0.3, 0.4, 0.7, 1.2*

*0.7 NS*) d) Fall time: *2.0 NS*

*1.2 NS*) e) PRF: *SEE TR*  
*0 TO 25 MHz.*

*2.0 NS*) f) Jitter, stability: *OK*

g) Prime power: *100 ± 10 VOLTS*  
*50-60 Hz.*

*40 dB ATTEN*

*5 VOLTS/DIV*

*PRF = 10 MHz.*