


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MX-D-3-PS*

S.N.: *0601*

Date: *OCT 20 1998*



- a) Output signal amplitude: *+3 V TO 50 V*
- b) Pulse width: *250 NS.*
- c) DELAY RANGE: *1 μS TO 100 μS.*
- d) PRF: *0 TO 50 KHz.*
- f) Jitter, stability: *✓*
- g) Prime power: *170/240 V*  
*50-60 Hz*

*DELAY ≈ 9 μS.*