


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MH-5-1-P*

S.N.: *8607*

Date: *OCT 14 1998*



- a) Output signal amplitude: *0 to +10 Volts (50 Ω)*
- b) Pulse width: *≤ 130 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*

40dB ATTEN

10 V/DIV

TOP 1.0 NS/DIV

B9I 100 ps/DIV

PRF = 300 kHz

150 mA (MAX)