


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AKR-EB-W-PS-F*

S.N.: *8609*

Date: *SEPT 28 1998*



- a) Output signal amplitude: *0 TO 700 V (R<sub>L</sub> = 50Ω)*
- b) Pulse width: *50 NS TO 5.0 μS*
- c) Rise time: *15 KHz MAX) ≤ 0.5 NS*
- d) Fall time: *≤ 1.0 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*  
*50-60 Hz*

*[Signature]*

*60 dB VOLTAGE*  
*100 V/DIV*  
*TOP 50 NS/DIV*  
*BOT 1 NS/DIV (RISE TIME)*