



PULSE GENERATOR
PERFORMANCE CHECK

Model: *NR14-1-C-P-011*
 S.N.: *8631*
 Date: *SEPT 3 1998*



- a) Output signal amplitude: *0 to +1000 V (R_L > 1k)*
- b) Pulse width: *0.5 to 5.0 μs*
- c) Rise time: *≤ 50 ns*
- d) Fall time: *≤ 50 ns*
- e) PRF: *0 to 1 kHz*
- f) Jitter, stability: *0%*
- g) Prime power: *120/248 V*
50-60 Hz

500 V / DIV
1 μs / DIV
R_i = 10k



[Signature]