


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-PS-P*

S.N.: *8639*

Date: *SEPT 2 1998*



- a) Output signal amplitude: *0 TO +350V* ($R_L = 50\Omega$)
- b) Pulse width: *5 TO 100 ns*
- c) Rise time: $\leq 2 ns$
- d) Fall time: $\leq 4 ns$
- e) PRF: *0 TO 5 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

60 db ATTN

100 VOLT/DIV

TOP: 10 ns/DIV

BOT: 2 ns/DIV
(RISE TIME)