


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MU-S-HVI-P*

S.N.: *8668*

Date: *DEC 10 1998*



- a) Output signal amplitude: *0 TO +40V*
- b) Pulse width: *≤ 350 PS (FWHM)*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*  
*200 mA (MAX)*

*[Signature]*

*40 dB ATTEN*  
*10 VOLTS/DIV*  
*PRF = 300 kHz*