


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-S-1A1-P*

S.N.: *3669*

Date: *DEC 10 1998*



- a) Output signal amplitude: *0 TO +40 V*
- b) Pulse width: *≤ 350 ps (FWHM)*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*  
*200 mA (max)*

*40 dB ATTEN*  
*10 VOLT/ DIV*  
*PRF ≈ 300 kHz*