


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AM-2-C-N-TN-CA*

S.N.: *2671*

Date: *DEC 9 1998*




- a) Output signal amplitude: *0 TO -350 V*
- b) Pulse width: *(50 ns)*
- c) Rise time: *5 TO 100 ns*
- d) Fall time: *≤ 2 ns*
- e) PRF: *≤ 3 ns*
- f) Jitter, stability: *OK*
- g) Prime power: *0 TO 5 kHz*

60 db ATTEN

100 V/V

PORT 2



← D - 60 dB

[Signature]

← 1703