


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV-110B-PS-PG1

S.N.: 8682

Date: DEC 17 1998



- a) Output signal amplitude:  $\pm 200 \text{ VOLTS (RL} \geq 50\text{K)}$
- b) GAIN:  $\times 1 \text{ TO } \times 100$
- c) Rise time:  $\leq 1 \mu\text{s}$
- d) Fall time:  $\leq 1 \mu\text{s}$
- e) PRF:  $\text{DC TO } 100 \text{ KHz}$
- f) Jitter, stability: OK
- g) Prime power:

120/240 V  
50-60 Hz

IN 2 V/DIV

OUT 50 V/DIV

$R_L = 100 \text{ K}$

5  $\mu\text{s/DIV (100 KHz)}$