


original photo
taken for use in
cat. 11.

PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMH-3-C-P-LIA*

S.N.: *8639*

Date: *FEB 4 1999*



- a) Output signal amplitude:
0 to +3000 V ($R_L \geq 10k$)
- b) Pulse width:
0.25 to 2.5 μ s
- c) Rise time:
 ≤ 100 ns
- d) Fall time:
 ≤ 100 ns
- e) PRF: *0 to 3 kHz.*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 J
50-60 Hz.



[Signature]