


PULSE GENERATOR

PERFORMANCE CHECK

MODEL: *ARF-4-PS-P*

S.N.: *8697*

DATE: *JAN 20 1999*



- A) Output signal amplitude: *0 TO +400 VOLTS*
- B) Pulse width: *(TO 50ns)*
- C) Rise time: *≤ 5ns*
- D) Fall time: *≤ 5ns*
- E) PRF: *0 TO 1KHz*

F) Jitter, stability: *OK*

G) Prime power: *120/240V*  
*50-60Hz*

*60 dB ATTEN*  
*- 100 VOLTS/DIV*  
*TOP 50 NS/DIV*  
*BUT 5 NS/DIV*  
*(RISE TIME)*  
*PRF ≈ 1 KHz*