


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AU-14V3-P-05*

S.N.: *8724*

Date: *JAN 7 1997*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width: *(50%)*
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

50 dB ATTEN
32 VOLTS/DIV
PRF ≈ 300 KHz

+15 V DC
300 mA MAX