


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-14V3-P-05*

S.N.: *8725*

Date: *JAN 7 1997*



- a) Output signal amplitude:  
*0 TO +40 VOLTS*
- b) Pulse width: *(50 ns)*  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 ps*
- d) Fall time:  
*≤ 250 ps*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 dB ATTEN*  
*32 VOLTS/DIV*  
*PRF ≈ 300 KHz*

*+15 V DC*  
*300 mA MAX*

*[Signature]*