


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMP-2-P*

S.N.: *8806*

Date: *MAR 9 1999*



- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *(50N)*
5 TO 100N)
- c) Rise time: *≤ 100 PS*
- d) Fall time: *6135 P*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *O.K.*
- g) Prime power:

+24 VOLTS

300 mA

MAX.

40 db ATTEN
10 VOLTS/DIV

TOP 10 NS/DIV

MRID 200PS/DIV (RISE TIME)

MRIF 200PS/DIV (FALL TIME)

PRF = 300 kHz