


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2A-PN-B*

S.N.: *3821*

Date: *APRIL 14 1999*



- a) Output signal amplitude:
0 TO ± 170 VOLTS
- b) Pulse width: *(TO 50 OHMS)*
3 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 5 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60 Hz

[Signature]

60db ATTN
... A 100 VOLTS/DIV
DVP Pout, 10 NS/DIV
MID Nout, 10 NS/DIV
BOT Pout, 2 NS/DIV
(RISE TIME)