


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUG-4A*
S.N.: *8846*
Date: *MAY 4 1999*



- a) Output signal amplitude:
+250 TO +650 VOLTS
- b) Pulse width: *(TO 50N)*
≤ 5 NS (AT 20% AMP)
- c) Rise time:
≤ 1.5 NS
- d) Fall time:
≤ 3.0 NS
- e) PRF: *TO 10 K C/P*
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
600 mA (MAX)

[Signature]